

TLN

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010488 Worksheet Start Date 5/14/2015

Location Northumberland Co SO

Address 76 Judicial Place Heathsville VA 22473

DFS Technician Timmy Neece, Jr License No. 25387

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 772 Reference Barometer (mm HG) 769

Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.311
0.300	0.305	0.295	0.314	Sample 2	0.311
Precision		sample min	sample max	Sample 3	0.311
0		0.311	0.311		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.102
0.100	0.102	0.098	0.105	Sample 2	0.102
Precision		sample min	sample max	Sample 3	0.102
0		0.102	0.102		

Dry gas standard Lot No. (with tank no.) AG406301-26

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Unable to call up instrument due to phone line disconnection for location change.

Instrument Serial Number

010488

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

5/14/15

Issuing Analyst



Date

5/14/15

JPD

TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED	NAME OF COURT
INSTRUMENT, TEST,	DFS

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY
NEECE, JR., TIMMY, L.		DFS Central Lab
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED
25387	10/01/2016	05/14/2015
TEST EQUIPMENT NUMBER		
010488		

RESULTS: TIME SAMPLE TAKEN 10:33 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 14 May 2015 at 11:34

Test Results

Instrument Serial Number 010488

Test # 002611 Subject Test

Test Location 1 Department of
Test Date 14 May 2015

Test Location 2 Forensic Science
Test Time 10:27
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE, JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

Effective Date 10/01/2014

License Number 25387

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 14 May 2015 End Time 10:34

Result Time 10:33

Result Date 14 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:27

Data Type BLK

Sample Value 0.000

Sample Time 10:28

Data Type CHK

Sample Value 0.102

Sample Time 10:29

Data Type BLK

Sample Value 0.000

Sample Time 10:30

Data Type SUBJ

Sample Value 0.000

Sample Time 10:31

Data Type BLK

Sample Value 0.000

Sample Time 10:32

Data Type SUBJ

Sample Value 0.000

Sample Time 10:33

Data Type BLK

Sample Value 0.000

Sample Time 10:34

Standard Type Dry Gas Std

Standard Value 0.102

Standard Lot Number AG406301-26

Standard Expiration Date 03/04/2016

Tank Pressure 321

Barometric Pressure 772 mmHg

Blow Sample Number 1 Blow Duration 3.30 sec

Blow Volume 1911 cc End-of-Blow Time 10:31

Blow Sample Number 2 Blow Duration 3.41 sec

Blow Volume 1782 cc End-of-Blow Time 10:33

Tamper Evident Stamp 343b254f

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

TL✓

Department of Forensic Science

Serial Number: 010488 Test Number: 2612

Test Date: 05/14/2015 Test Time: 10:36 EDT

Dry Gas Target: 0.102

Lot Number: AG406301-26 Exp Date: 03/04/2016

Tank Pressure: 321 psi Barometric Pressure: 772 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:37
CHK	0.102	10:37
BLK	0.000	10:39
CHK	0.102	10:40
BLK	0.000	10:41
CHK	0.102	10:42

Test Status: *Success*

Calibration CRC: AE27756B

Intox EC/IR-II: Accuracy Check

+L✓

Department of Forensic Science

Serial Number: 010488 Test Number: 2613

Test Date: 05/14/2015 Test Time: 10:46 EDT

Dry Gas Target: 0.305

Lot Number: AG428003-02 Exp Date: 10/07/2016

Tank Pressure: 267 psi Barometric Pressure: 772 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	10:47
CHK	0.311	10:47
BLK	0.000	10:49
CHK	0.311	10:50
BLK	0.000	10:52
CHK	0.311	10:53

Test Status: Success

Calibration CRC: AE27756B